

CE

## Declaration of Conformity

*Manufacturer:*

*Name: Henan Bochuang Medical Technology Co., Ltd*

*Add: North Ring Road of Changyuan city and West 1000m north of provincial highway S308*

*Product Name : Medical isolation shield*

*Risk Class of the Device: The medical device has been assigned to class I , rule 1 according to ANNEX VIII, Medical Device REGULATION (EU) 2017/745.*

*As a manufacturer, we declare that  
The product concerned has been designed and manufactured under a quality management system according to ANNEX IX of Medical Device REGULATION (EU) 2017/745.*

*Following the procedure relating to the EC Declaration of Conformity set out in ANNEX IX of Medical Device REGULATION (EU) 2017/745.*

*This Declaration of conformity is valid in connection with the release document for the respective batch of produced devices*

*European Representative:*

*Name: Llins Service & Consulting GmbH*

*Address: Obere Seegasse 34/2, 69124,*

*Heidelberg, Germany*

*DIMDI Code: DE/0000048234*

*Tel: +49 175 4870 819*

*Email: info@llins-service.com*

*Manufacturer:*

*Gao chao*

*date: 2020.03.19*